



# UNITED STATES PATENT AND TRADEMARK OFFICE

**UNITED STATES DEPARTMENT OF COMMERCE**  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

A

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/627,313	07/24/2003	Philip E. Eggers	A-3-4	1924

21394 7590 04/07/2005  
ARTHROCARE CORPORATION  
680 VAQUEROS AVENUE  
SUNNYVALE, CA 94085-3523

EXAMINER

THOMPSON, MICHAEL M

ART UNIT PAPER NUMBER

3763

DATE MAILED: 04/07/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

**Application No.**

10/627,313

**Applicant(s)**

EGGERS ET AL.

**Examiner**

Michael M. Thompson

**Art Unit**

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 01/14/2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 54-66 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 54-66 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |                                                                                                                                               |                                                                                         |
|-----------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                                                                              | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                                          | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date <u>2/7/2005</u> . | 6) <input type="checkbox"/> Other: _____                                                |

## **DETAILED ACTION**

### ***Information Disclosure Statement***

1. The listing of references in the specification is not a proper information disclosure statement. 37 CFR 1.98(b) requires a list of all patents, publications, or other information submitted for consideration by the Office, and MPEP § 609 A(1) states, "the list may not be incorporated into the specification but must be submitted in a separate paper." Therefore, unless the references have been cited by the examiner on form PTO-892, they have not been considered.

### ***Specification***

2. The lengthy specification has not been checked to the extent necessary to determine the presence of all possible minor errors. Applicant's cooperation is requested in correcting any errors of which applicant may become aware in the specification.

### ***Claim Rejections - 35 USC § 102***

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

4. Claims 54-56 and 65-66 are rejected under 35 U.S.C. 102(b) as being anticipated by Roos (4,116,198). Roos teaches a surgical instrument for applying high frequency electrical energy to tissue at a target site comprising a shaft, a hemispherical-shaped electrode terminal (12), an annular return electrode (11) spaced proximally from the hemispherical-shaped electrode, a

Art Unit: 3763

connector (14,16) extending from the electrode terminal to the proximal end of the shaft and an electrode support (14,22).

*Claim Rejections - 35 USC § 103*

5. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

6. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

7. Claims 57-58 and 60-64 are rejected under 35 U.S.C. 103(a) as being unpatentable over Roos in view of Herczog et al. (GB 2037167). Roos teaches all of the limitations of the claims except for explicitly reciting that the electrode support is glass or ceramic or a bent shaft configuration. Herczog et al. teaches the use of glass and ceramic as an insulating electrode support and shows a bent shaft configuration. It would have been obvious to one of ordinary skill in the art, at the time of invention, to modify the insulating electrode support of Roos with the glass and ceramic insulating electrode support and bend taught by Herczog et al. for the well

Art Unit: 3763

known purpose of providing proper insulation to electrical/ high frequency applications and providing for a transverse treatment location for structures parallel to the device.

### ***Double Patenting***

8. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

9. Claims 54-66 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-59 of U.S. Patent No. 5,697,909. Although the conflicting claims are not identical, they are not patentably distinct from each other because both substantially recite a surgical instrument for applying high frequency electrical energy to tissue at a target site comprising a shaft, an electrode terminal, an annular return electrode spaced proximally from the hemispherical-shaped electrode, a connector extending from the electrode terminal to the proximal end of the shaft and an electrode support.

### ***Response to Arguments***

Applicant's arguments filed 01/14/2005 have been fully considered but they are not persuasive.

The main thrust of Applicant's arguments asserts that the Roos Patent fails to teach a active electrode with a "hemispherical geometry." Applicant further states that the electrode (12) of

Art Unit: 3763

Roos is merely "U-shaped." The Examiner respectfully disagrees with Applicant's assertion that electrode (12) of Roos fails to constitute a hemispherical geometry. On the outset, it should be noted that Merriam-Webster's Online Dictionary, 10th Edition has defined the term hemisphere as, "Hemisphere: half of a spherical or roughly spherical body." It is the Examiner's position that at least a portion of electrode (12) depicted in Roos is well within the definition of a hemisphere. Furthermore, the Examiner would like to direct Applicant's attention to Figure 1 which depicts a generally hemispherical/spherical shape as the scope of Roos patented invention. In conclusion, it is understood that Applicant has fully responded to all rejections and raised all errors with respect to the Examiner's rejection. Since there is no further traversal beyond the above-mentioned issues it appears that Applicant has acquiesced to all other rejections of record and therefore the instant Office Action has been made Final.

### ***Conclusion***

10. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Art Unit: 3763

### Contacts


Any inquiry concerning this communication or earlier communications from the Examiner should be directed to Michael Thompson whose telephone number is (571) 272-4968.

The Examiner can normally be reached on Monday through Friday from 9 am to 5 PM.

If attempts to reach the Examiner by telephone are unsuccessful, the Examiner's Supervisor, Nick Lucchesi, can be reached on (571) 272-4977. The official fax phone number for all submissions to the organization where this application or proceeding is assigned is (703) 872-9306.

Michael M. Thompson

Patent Examiner



NICHOLAS D. LUCCHESI  
SUPERVISOR/PATENT EXAMINER  
TECHNOLOGY CENTER 3700

MT



March 22, 2005